Finished product inspection report

Serial number: IR/QP-12-F05 A/O

Serial number:	JR/QP-12-F05 A/0						
Product name	Self sealing pouch	Prod	uction order		W07AMT2301		
Model number	/	Spe	cifications		57*130mm		
Lot number	/	Temp	perature and		28.5℃/55 RH%		
Lot qty	280000	,	Sampling		80		
Date of	2023. 6. 27	Dat	e of report		2023. 6. 30		
Inspection	SIP-19 《Standard for inspection of single-use sterilized bags》						
Inspection	Inspection Standar	rd			Test results	Concl usion	
Visual inspection	1 The sterilizing bag shall not have any oil or foreign matter which				It fits the bill	Quali	
	obviously affects the appearance or cleanliness. 2 hold the flat sterilizing bag with both hands to seal the edge				It fits the bill		
	without obvious wave phenomenon.						
	3 printing should not have the Burr, the air print affects the recognition effect.				It fits the bill		
	4 Printing surface should not be stained or dragged when making bags.				It fits the bill		
	5 Packing method and marking shall be in accordance with factory standard and customer requirements.				It fits the bill		
	6 The quantity (volume/box) shall be in accordance with the packing indication and customer requirements.			It fits the bill			
	7 The indicator color shall conform to the color card before and after sterilization.				It fits the bill		
	The surface, dimension, structure and mate	100					
	of the disinfection bag should be consistent		Bag Width		56-58mm	1	
Inspection of specifications					9-11mm		
	with that indicated on the surface, the length Sealing width		9 11111111	Quali			
	and width error is \pm 2 mm, the sealing width and the thumb opening error is \pm 1 mm		14-16mm				
	Printing surface weight error ± 5%	60g Arjo paper		$58-61 \text{g/m}^2$	fied		
	Weight/thickness error \pm 5%	5.2s high temperature film e		0.051-0.053mm			
	The printed area of the sterilization indication is $\geqslant \! 100 \text{ mm}^2$			It fits the bill			
Seal Strength Test	The sample is cut 15mm in width and		Before sterilization≥ 2.1N/15mm		2.5-3.3N/15mm	Quali fied	
	pulled apart at a speed of 200mm/min	speed of 200mm/min After sterilization≥ 1.5N/15mm		ion≥	1.8-2.3N/15mm	Quali fied	
Sealing edge leakage test	A syringe (which can be made of plastic or glass) is used to absorb an appropriate amount of aqueous solution of dye and inject it into the disinfection bag. After the injection of Dye, the depth of the water solution on the sealing edge of the disinfection bag should be greater than 5 mm, the Dye is evenly distributed on one side of the sterilizing bag and left to rest for 5-20 seconds. If there is any leakage of 1/2 width of the sealing specification, it is regarded as				It fits the bill	Quali fied	
	unqualified						

Tianchang Jiarui Medical Materials Co., Ltd.

Tearing test	测Pull back the fastening point of the bottom seal, slowly and carefully open the sterilizing bag with both hands, apply force to two layers of material and tear apart the two sides. On the sealing edge, can not produce more than 10 mm of paper fracture	It fits the bill	Quali fied
Tolerance test for disinfection	The sterile bag is divided into two parts. One part is tested and the other part is sealed with gauze or the corresponding product. After sterilization, the bag should not be broken at any sealing place. And printing can not produce halo-dye or decolorization, composite film melt crack and so on, to observe the composite film can not have two layers of peeling off the film	It fits the bill	Quali fied

CONCLUSION: This product is tested according to SIP-19 《Standard for inspection of single-use sterilized bags》, and the test results of this batch of products are qualified.

天长市嘉瑞医用材料有限公司 TIANCHANG JIARUI PACKAGING MATERIAL CO., LTD.

Date: 2023.6.30

Test: FEI. WANG Recheck: YUANFANG. TAO

Permission granted: WEN. CHENG